



Product Specification
**UriSet 24 NFT, with stabiliser, 3 l, with
Urine-Monovette® 10 ml, brown, with light
protection, graduated**

Page 1

Product description

Order number	77.578.952
Product description	UriSet 24 NFT, Complete set for optimum 24-hour urine collection, max. working volume: 3 l, Ø opening: 83.5 mm, with stabiliser, brown, with light protection, graduated, material: PE, cap: white, with Urine-Monovette® 10 ml, 30 piece(s)/case

Product characteristics

Type of preparation	with stabiliser
Graduation	yes
Cap	screw cap with needle-free transfer device
Closure type	screw cap
Light protection	yes
Design	with Urine-Monovette® 10 ml
Application	Complete set for optimum 24-hour urine collection
Viewing window	yes

Size

Sample volume	3.000.00 ml
Max. volume	3 l
Diameter of opening	83,5 mm
Width of product	130 mm
Length of product	125 mm
Length including cap	246 mm
Length excluding cap	243 mm

Material & colours

Product material	Polyethylene (PE)
Colour of product	brown
Closure material	Polypropylene (PP)
Colour of cap	white

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

Date of issue:
2024-10-19

This document was prepared by EDP support and is valid without signature.



SARSTEDT



Product Specification
UriSet 24 NFT, with stabiliser, 3 l, with
Urine-Monovette® 10 ml, brown, with light
protection, graduated

Page 2

Purity & certification

Product category	In vitrodiagnostic device, CE
CE certified	CE
Batched	yes

Packaging

Minimum order qty.	30
Type of smallest subpackaging	case
Piece(s) / inner box	30
Piece (s) / outer case	30
Piece(s) / pallet	270
Depth of case	643 mm
Width of case	403 mm
Height of case	515 mm
Case volume	0.1335 cbm
Weight of product	0.2296 kg
Weight of case	7.69 kg
EAN of inner box	4038917983228
EAN case	4038917983235

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

Date of issue:
2024-10-19

This document was prepared by EDP support and is valid without signature.



SARSTEDT