




Product Specification

Placa PCR sem borda, 96 poço, transparente,
Perfil baixo, 100 µl, PCR Performance Tested, PP

Page 1



Product description

Order number	72.1977.202
Descrição do artigo	Placa PCR sem borda, 96 poço, transparente, Perfil baixo, 100 µl, PCR Performance Tested, material: PP, 20 unid./pacote com minialça

Product characteristics

Tipo de fechamento	filme
Indicada para qPCR	sim
Forma do fundo	cônico
Forma da borda da placa	sem borda
High profile ou Low profile	Perfil baixo
Quantidade de poços	96
Volume de trabalho	5 µl - 0,1 ml
Centrifugação máxima (RCF)	3000 x g
Temperatura de centrifugação	22
Tempo de centrifugação	5 Min
Autoclavável (máx. °C)	121 °C

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

Date of issue:
2024-07-18

This document was prepared by EDP support and is valid without signature.



SARSTEDT



Product Specification

**Placa PCR sem borda, 96 poço, transparente,
Perfil baixo, 100 µl, PCR Performance Tested, PP**

Page 2

Size

Volume nominal	0.1 ml
Volume máximo	100 µl
Largura do produto	77.7 mm
Comprimento do produto	117.9 mm

Material & colours

Material do produto	Polipropileno (PP)
Cor do produto	transparente

Purity & certification

Categoria do produto	Não é um dispositivo médico DIV
Padrão de pureza	PCR Performance Tested
Em lotes	sim

Packaging

Quantidade mínima do pedido	100
Tipo de embalagem interior	pacote com minialça
Unidades por caixa interna	20
Unidades por caixa	100
Unidades por palete	5000
Profundidade da embalagem externa	137 mm
Largura da embalagem externa	137 mm
Altura da embalagem externa	493 mm
Volume da embalagem externa	0.0093 cbm
Peso do produto	0.0128 kg
Peso da embalagem externa	1.5 kg
EAN da caixa interna	4038917405102
EAN da embalagem externa	4038917405096

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

Date of issue:
2024-07-18

This document was prepared by EDP support and is valid without signature.



SARSTEDT